

Product Introduction

- Model:

F-Y1-A Surgical face mask

- Style : Flat
- Wearing style: Earloop
- Exhalation Valve : None
- Filter Level: BFE 98
- Color blue
- Material :

Surface layer - 25g non woven fabric

- Second layer - 20g BFE99 filter material
- Inner layer - 25g non woven fabric
- Certification : FDA CE



Test report



检 验 报 告

Test Report

报告编号: Z20200622

产 品 名 称: 医用外科口罩

规 格 型 号: F-Y1-A

受 检 单 位: 建德市朝美日化有限公司

检 验 类 别: 委托检验

浙江省医疗器械检验研究院
ZHEJIANG INSTITUTE OF MEDICAL DEVICE SUPERVISION AND TESTING

浙江省医疗器械检验研究院

检 验 报 告

报告编号: Z20200622

共 5 页 第 1 页

样品名称	医用外科口罩		
规格型号	F-Y1-A	检验类别	委托检验
委托人/单位	建德市朝美日化有限公司		
受检单位名称	建德市朝美日化有限公司		
制造单位名称	建德市朝美日化有限公司		
取样方式	送样	抽样地点	/
抽样日期	/	抽样基数	/
抽样人	/	样品接受日期	2020-03-05
样品数量	50只	样品生产日期	2020.02.15
样品批号/编号	20200215		
检验依据	YY 0469-2011		
检验项目	全项目(除4.9.2、4.10、4.11、4.12、4.13)		
检验日期	2020年03月05日~2020年03月15日		
检验结论	被检样品所检项目YY 0469-2011《医用外科口罩》的要求。		

YY0469-2011 China standard



批准: 张莉 职务: 授权签字人 日期: 2020.03.31
 审核: 周明 日期: 2020.03.30
 主检: 王丽群 日期: 2020.03.27

Nelson Test report



Sponsor:
Rolen He
Jiande Chaomei Daily Chemicals Co.,Ltd.
Shangshan village Yangcungiao town
Jiande city, Zhejiang Province, 311600
CHINA



Study Number 1293322-S01
Latex Particle Challenge Final Report

Latex Particle Challenge Final Report

Test Article: MEDICAL SURGICAL MASKS /Model: F-Y1-A
Purchase Order: 20-243A
Study Number: 1293322-S01
Study Received Date: 27 Apr 2020
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd.
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP0005 Rev 07
Deviation(s): Quality Event (QE) Number(s): QE22125

Summary: This procedure was performed to evaluate the non-viable particle filtration efficiency (PFE) of the test article. Monodispersed polystyrene latex spheres (PSL) were nebulized (atomized), dried, and passed through the test article. The particles that passed through the test article were enumerated using a laser particle counter.

A one-minute count was performed, with the test article in the system. A one-minute control count was performed, without a test article in the system, before and after each test article and the counts were averaged. Control counts were performed to determine the average number of particles delivered to the test article. The filtration efficiency was calculated using the number of particles penetrating the test article compared to the average of the control values.

The procedure employed the basic particle filtration method described in ASTM F2299, with some exceptions; notably the procedure incorporated a non-neutralized challenge. In real use, particles carry a charge, thus this challenge represents a more natural state. The non-neutralized aerosol is also specified in the FDA guidance document on surgical face masks. All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Side: Inside
Area Tested: 91.5 cm²
Particle Size: 0.1 µm
Laboratory Conditions: 21°C, 23% relative humidity (RH) at 1042; 21°C, 23% RH at 1139
Average Filtration Efficiency: 99.87%
Standard Deviation: 0.048



Christopher Acker electronically approved for
Study Director

Curtis Gerow

10 Jun 2020 05:42 (+00:00)
Study Completion Date and Time

Deviation Details: Controls and sample counts were conducted for one minute instead of an average of three one minute counts. This change shortens the total test time for each sample but will still provide an accurate determination of the particle counts. An equilibrate is a dwell period where the challenge is being applied to the test article for a certain period of time before test article counts are counted. The equilibrate period was reduced from 2 minutes to a minimum of 30 seconds which is sufficient time to clear the system of any residual particles, and establish a state of stable equilibrium before sample counts are taken. Test method acceptance criteria were met, results are valid.

Results:

Test Article Number	Test Article Counts	Average Control Counts	Filtration Efficiency (%)
1	49	13,587	99.84
2	50	13,602	99.83
3	40	13,543	99.70
4	36	13,443	99.73
5	48	12,897	99.83

BFE ≥ 99%

Nelson Test report



Sponsor:
Rolen He
Jande Chaomei Daily Chemicals Co.,Ltd.
Shangshan Village Yangcunqiao Town
Jande City, Zhejiang Province
CHINA

Synthetic Blood Penetration Resistance Final Report

Test Article: MEDICAL SURGICAL MASKS /Model: F-Y1-A
Purchase Order: 20-243A
Study Number: 1293323-S01
Study Received Date: 28 Apr 2020
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd.
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP0012 Rev 09
Deviation(s): None

Summary: This procedure was performed to evaluate surgical facemasks and other types of protective clothing materials designed to protect against fluid penetration. The purpose of this procedure is to simulate an arterial spray and evaluate the effectiveness of the test article in protecting the user from possible exposure to blood and other body fluids. The distance from the target area surface to the tip of the cannula is 30.5 cm. A test volume of 2 mL of synthetic blood was employed using the targeting plate method.

This test method was designed to comply with ASTM F1862 and ISO 22809 (as referenced in EN 14683:2019 and AS4381:2015) with the following exception: ISO 22809 requires testing to be performed in an environment with a temperature of $21 \pm 5^\circ\text{C}$ and a relative humidity of $85 \pm 10\%$. Instead, testing was performed at ambient conditions within one minute of removal from the environmental chamber held at those parameters.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Number of Test Articles Tested: 32
Number of Test Articles Passed: 31
Test Side: Outside
Pre-Conditioning: Minimum of 4 hours at $21 \pm 5^\circ\text{C}$ and $85 \pm 5\%$ relative humidity (RH)
Test Conditions: 20.4°C and 22% RH

Results: Per ASTM F1862 and ISO 22809, an acceptable quality limit of 4.0% is met for a normal single sampling plan when ≥ 29 of 32 test articles show passing results.

Test Pressure: 120 mmHg (16.0 kPa)

Test Article Number	Synthetic Blood Penetration
1-25, 27-32	None Seen
26	Yes



Sean Shepherd electronically approved for
Study Director James Luskin

09 May 2020 19:15 (+00:00)
Study Completion Date and Time

Level:2
Test Pressure:
120 mmHg (16.0 kPa)

FDA 510K & Listing


Fiscal Year 2020
CERTIFICATION OF FDA REGISTRATION

Cert. No.: 263DF2AC85F0CAC8042

This certifies that:

Establishment: JIANDI CHAOMEI DAILY CHEMICALS CO., LTD
Registered Address: Changshu village Yangcunqiao town Jiande, Zhejiang, 311801, CHINA
Registration Number: Not Yet Assigned

has completed the FDA Establishment Registration (as manufacturer and foreign exporter) and Device Listing with the US Food & Drug Administration, such registration having been verified as currently effective on the date hereof by SPICA MEDTECH CORP.

Owner/Operator Number: 10062722 (Device Listing #: See above.)
U.S. Agent for FDA: SPICA MEDTECH CORP
Communications: 3285 KNOX DR Denver, CO 80205 USA
Phone: 720 6176666 Ext. Email: spica_us@yahoo.com

SPICA MEDTECH CORP will continue that such registration remains effective upon request and presentation of this certificate until the end of the calendar year stated above, unless said registration is terminated after issuance of this certificate. SPICA MEDTECH CORP makes no other representations or warranties, nor does this certificate make any representations or warranties in any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by the U.S. Food and Drug Administration.

Pursuant to 21 CFR 807.59, "Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products."

SPICA MEDTECH CORP (in USA)
SHANGHAI SPICA MANAGEMENT CONSULTING CO., LTD (in China)
SPICA MEDTECH CERTIFICATION COMPANY LIMITED (in HK)
spica_us@yahoo.com, www.spicamed.com



For and on behalf of
SPICA MEDTECH CORP
Silverlan Wu
Authorized Signatory

 

SPICOR CORP 499 No. 231 Building No. 88 Daxin Road, Shanghai 201201, P.R. China

Follow FDA | En Español

FDA U.S. FOOD & DRUG ADMINISTRATION

[Home](#) | [Food](#) | [Drugs](#) | [Medical Devices](#) | [Radiation-Emitting Products](#) | [Vaccines, Blood & Biologics](#) | [Animal & Veterinary](#) | [Cosmetics](#) | [Tobacco Products](#)

Establishment Registration & Device Listing

[FDA Home](#) | [Medical Devices](#) | [Databases](#)

  

New Search	Back To Search Results
Proprietary Name: F-Y1-A MEDICAL SURGICAL MASKS; F-Y3-A N95 medical protective mask; Y3-A N95 medical protective mask	
Classification Name: MASK, SURGICAL	
Product Code: FXX	
Device Class: 2	
Regulation Number: 873.4040	
Medical Specialty: General & Plastic Surgery	
Registered Establishment Name: JIANDI CHAOMEI DAILY CHEMICALS CO., LTD	
Registered Establishment Number: 3010610938	
Premarket Submission Number: K142990	
Owner/Operator: Jiandi Chaomei Daily Chemicals Co. Ltd	
Owner/Operator Number: 10062722	
Establishment Operations: Contract Manufacturer, Foreign Exporter, Manufacturer	

Page Last Updated: 05/11/2020
Note: If you need help accessing information in different file formats, see instructions for Downloading Viewers and Players.
Language Assistance Available: Español | 繁體中文 | Tiếng Việt | 한국어 | Tagalog | Pycckий | العربية | Kreyòl Ayisyen | Français | Polski | Português | Italiano | Deutsch | 日本語 | العربية | English

D40094 K142990 FXX

FDA 510K & Listing



DEVICE REGISTRATION AND LISTING INFORMATION

Registration Date: May 27, 2020

Registration Status: Active

Please review your registration.

Section 01		Type of Registration
Activities	MANUFACTURER, CONTRACT MANUFACTURER, FOREIGN EXPORTER	
Registration Number	3016610938	

Section 02		Facility Name / Address Information
Name	JIANDE CHAOMEI DAILY CHEMICALS CO.,LTD.	
Address	Shangshan village Yangcunqiao town, Jiande, Zhejiang, 311603, CHINA	

Section 03		Owner/Operator Information
Owner/Operator Number	10062722	
Contact Name	Rolen He	
Company	JIANDE CHAOMEI DAILY CHEMICALS CO.,LTD.	
Address	Shangshan village Yangcunqiao town, Jiande, Zhejiang, 311603, CHINA	
Phone Number	86 - 571 - 6419000 - 3	
Fax Number		
Email Address	Rolen@cmmask.com	

Section 04		Official Correspondent Information
OC Contact Name	Rolen He	
OC Business Name	JIANDE CHAOMEI DAILY CHEMICALS CO.,LTD.	
Address	Shangshan village Yangcunqiao town, Jiande, Zhejiang, 311603, CHINA	
Phone Number	86 - 571 - 6419000 - 3	
Fax Number		
Email Address	Rolen@cmmask.com	

Section 05		Trade Names

Section 06		United States Agent
Contact Name	SILVERDEW VEE	
Business Name	SPICA MEDTECH CORP	
Address	2255 EMRSON ST DENVER, COLORADO 80205 UNITED STATES	

Contact Name:
Rolen He

Company:
JIANDE CHAOMEI DAILY CHEMICALS CO.,LTD.

Address:
Shangshan village Yangcunqiao town
jiande, ZHEJIANG, 311603, CHINA.

Telephone:
86 - 571 - 6419000 - 3

Fax:
-

E-mail:
Rolen@cmmask.com

DUNS Number:

United States Agent Information

Contact Name:
Silverdew Vee

Contact Title:
Mrs

Business Name:
SPICA MEDTECH CORP

Address:
2255 EMRSON ST
Denver, Colorado, 80205, UNITED STATES

Phone:
720 - 6176666

Fax:
720 - 6176666

DUNS Number:

E-mail:
Spica_us@yahoo.com

Device Listings

Listing Number	Premarket Submission Number/Type	Product Code(s)	Device Name(s)	Activities	Importers
D879361	Exempt	LYU	ACCESSORY, SURGICAL APPAREL	Foreign Exporter Manufacturer	
D400094	K142990	FXX	Mask, surgical	Manufacturer Contract Manufacturer Foreign Exporter	

https://www.access.fda.gov/fda/rdm.htm?_flowExecutionKey=_c480D1F40-F700-618E-BBF4-557590F89C3F_k79C1E885-1C4B-C1AA-6E8B-... 2/3

510K : K142990

Listing Number	Proprietary Name	Confidential Flag
D373391	Medical face mask	N
D400094	Y3-A N95 medical protective mask	N
D400094	F-Y1-A MEDICAL SURGICAL MASKS	N
D400094	F-Y3-A N95 medical protective mask	N

Packaging

50pcs /box 2000pcs/carton
Carton size: 53cm*40cm*36cm
Carton weight (G.W.): 9 kgs

